

## Certificate of Analysis

---

**Material:**      PROL2678.1000      **AMMONIUM CERIUM IV SULFATE 0,1  
MOL/L**

**Batch :**          133665      **Expires end of**      23/12/2014

---

<b>CHARACTERISTICS</b>	<b>SPECIFICATIONS</b>	<b>MESURED VALUES</b>
Titer (20°C; real value 0.2 % accuracy)	Specification= 0,1000 mol/l+/-0,2 %	0,1000 N

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

Signé par:



---

O. BRINGER , Responsable Qualité