

Certificate of Analysis

Material: 88946320 **Sodium thiosulfate 0,039 N**

Batch : 143732 **Expires end of** 12/06/2015

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,0394 N +/- 0,5%	0,03932 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

05/01/2015

Signé par:

O. BRINGER , Responsable Qualité