

Certificate of Analysis

Material: MC3078671000 **Acide nitrique 0,1 N**

Batch : 143845 **Expires end of** 23/12/2015

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,100 N +/- 0,2%	0,1002 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

05/01/2015

Signé par:

O. BRINGER , Responsable Qualité