

Certificate of Analysis

Material: MC3125572500 **Acide sulfurique 2,5 N**

Batch : 150164 **Expires end of** 19/01/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,500 N +/- 0,2%	2,500 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

26/01/2015

Signé par:

O. BRINGER , Responsable Qualité