

# Certificate of Analysis

---

**Material:** MC3023350500      **Sodium Hydroxyde 2 N**

**Batch :** 150275      **Expires end of** 27/01/2016

| CHARACTERISTICS                         | SPECIFICATIONS           | MESURED VALUES |
|-----------------------------------------|--------------------------|----------------|
| Titer (20°C; real value 0.2 % accuracy) | Titre = 2,000 N +/- 0,2% | 2,002 N        |

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**  
  
**05/02/2015**

**Signé par:**  
  
O. BRINGER , Responsable Qualité