

Certificate of Analysis

Material: 88946320 **Sodium thiosulfate 0,039 N**

Batch : 150430 **Expires end of** 10/08/2015

| <i>CHARACTERISTICS</i> | <i>SPECIFICATIONS</i> | <i>MESURED VALUES</i> |
|---|------------------------------|------------------------------|
| Titer (20°C; real value 0.2 % accuracy) | Titre = 0,0394 N +/- 0,5% | 0,03934 N |

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/02/2015

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité