

Certificate of Analysis

Material: MC3104741000 **Potassium hydroxyde 0,05 N**

Batch : 150442 **Expires end of** 11/02/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,05 N +/- 0,2%	0,0499 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/02/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité