

Certificate of Analysis

Material: MC3139131000 **ACIDE CHLORHYDRIQUE 1 N DANS
L'ETHYLENE GLYCOL/PROPANOL**

Batch : 150443 **Expires end of** 11/02/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2 %	1,002 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

23/02/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité