

Certificate of Analysis

Material: MC3107015010 **Acide chlorhydrique 4 N**

Batch : 150492 **Expires end of** 16/02/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 4,000 N +/- 0,2%	4,001 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

23/02/2015

Signé par:
A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.
O. BRINGER , Responsable Qualité