

Certificate of Analysis

Material: MC3023281000 **Potassium hydroxyde N solution alcoolique**

Batch : 150653 **Expires end of** 02/09/2015

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2%	0,999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

09/03/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité