

Certificate of Analysis

Material: MC3067695000 **Acide sulfurique 2 N**

Batch : 150782 **Expires end of** 12/03/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,000 N +/- 0,2%	1,998 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

16/03/2015

Signé par:

O. BRINGER , Responsable Qualité