

# Certificate of Analysis

---

**Material:** MC3078671000      **Acide nitrique 0,1 N**

**Batch :** 150925      **Expires end of** 25/03/2016

---


<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,100 N +/- 0,2%	0,1002 N

---

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**  
  
**30/03/2015**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité