

Certificate of Analysis

Material: MC3023361000 **Sodium hydroxyde 1 N**

Batch : 151122 **Expires end of** 13/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2 %	0,999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

24/04/2015

Signé par:

O. BRINGER , Responsable Qualité