

Certificate of Analysis

Material: 85125320 **SODIUM HYDROXYDE 2,5 N**

Batch : 151237 **Expires end of** 23/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,500 N +/- 0,2 %	2,498 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

04/05/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité