

Certificate of Analysis

Material: MC3093802500 **Acide chlorhydrique 6 N**

Batch : 151319 **Expires end of** 28/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 6,000 N +/- 0,2%	6,01 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

04/05/2015

Signé par:

O. BRINGER , Responsable Qualité