

Certificate of Analysis

Material: MC2403375000 **Titriplex 0,2 N**

Batch : 151328 **Expires end of** 29/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,200 N +/- 0,2%	0,1996 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

04/05/2015

Signé par:

O. BRINGER , Responsable Qualité