

# Certificate of Analysis

---

**Material:** MC3023360500      **Sodium hydroxyde 1 N**

**Batch :** 151338      **Expires end of** 30/04/2016

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2 %	0,999 N

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**  
  
**18/05/2015**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité