

Certificate of Analysis

Material: MC3052891000 **Sodium hydroxyde 0,01 N**

Batch : 151339 **Expires end of** 30/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,01000 N +/- 0,2%	0,01001 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

18/05/2015

Signé par:

O. BRINGER , Responsable Qualité