

Certificate of Analysis

Material: MC3074289010 **Acide sulfurique 2,6 N**

Batch : 151368 **Expires end of** 04/05/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,600 N +/- 0,2%	2,602 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

18/05/2015

Signé par:

O. BRINGER , Responsable Qualité