

Certificate of Analysis

Material: MC307586.460 **Acide chlorhydrique 0,04 N**

Batch : 151462 **Expires end of** 18/05/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titrage = 0,040 N +/- 0,2%	

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

18/05/2015

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité