

Certificate of Analysis

Material: MC3093819010 **Acide chlorhydrique 3 N**

Batch : 151606 **Expires end of** 02/06/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 3,000 N +/- 0,2%	3,003 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

08/06/2015

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité