

Certificate of Analysis

Material: MC3098066010 **Potassium hydroxyde 1 N**

Batch : 151776 **Expires end of** 12/06/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2%	0,999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

23/06/2015

Signé par:

O. BRINGER , Responsable Qualité