

Certificate of Analysis

Material: MC3052875000 **Argent nitrate 0,01 N**

Batch : 151806 **Expires end of** 16/06/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,01000 N +/- 0,2%	0,01000 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

23/06/2015

Signé par:

O. BRINGER , Responsable Qualité