

# Certificate of Analysis

---

**Material:** MC3067600500      **Acide sulfurique 0,1 N**

**Batch :** 151830      **Expires end of** 18/06/2016

| CHARACTERISTICS                         | SPECIFICATIONS           | MESURED VALUES |
|-----------------------------------------|--------------------------|----------------|
| Titer (20°C; real value 0.2 % accuracy) | Titre = 0,100 N +/- 0,2% | 0,1000 N       |

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**  
  
**23/06/2015**

**Signé par:**

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité