

Certificate of Analysis

Material: 85123290 **IODE SOLUTION 0,02 N**

Batch : 151949 **Expires end of** 29/06/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,02000 N +/- 0,5 %	0,01997 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

07/07/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité