

Certificate of Analysis

Material: MC3033631000 **Argent nitrate 0,2 N**

Batch : 152093 **Expires end of** 09/07/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,2000 N +/- 0,2%	0,1999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

21/07/2015

Signé par:

O. BRINGER , Responsable Qualité