

Certificate of Analysis

Material: MC3107015010 **Acide chlorhydrique 4 N**

Batch : 152164 **Expires end of** 20/07/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 4,000 N +/- 0,2%	3,998 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

27/07/2015

Signé par:

O. BRINGER , Responsable Qualité