

Certificate of Analysis

Material: MC3093811000 **Acide chlorhydrique 3 N**

Batch : 152421 **Expires end of** 14/08/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 3,000 N +/- 0,2%	3,003 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

02/09/2015

Signé par:

O. BRINGER , Responsable Qualité