

Certificate of Analysis

Material: 88885260 **TITRANT CHLORURES 1,128N**

Batch : 152519 **Expires end of** 25/08/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	TITRE=1,128N+/-0,2%	1,127 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

02/09/2015

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité