

Certificate of Analysis

Material: MC3096161000 **Sodium hydroxyde 0,05 N**

Batch : 152597 **Expires end of** 04/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,0500 N +/- 0,2%	0,0510 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2015

Signé par:

O. BRINGER , Responsable Qualité