

Certificate of Analysis

Material: BD19168xx **Acide sulfurique 2 N**

Batch : 152603 **Expires end of** 04/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,000 N +/- 0,2%	1,999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

07/09/2015

Signé par:

O. BRINGER , Responsable Qualité