

Certificate of Analysis

Material: MC3067875000 **Acide nitrique 0,5 N**

Batch : 152616 **Expires end of** 07/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,500 N +/- 0,2%	0,499 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2015

Signé par:

O. BRINGER , Responsable Qualité