

Certificate of Analysis

Material: MC3023316010 **Sodium thiosulfate 0,1 N**

Batch : 152637 **Expires end of** 08/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,100 N +/- 0,2%	0,0999 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2015

Signé par:

O. BRINGER , Responsable Qualité