

Certificate of Analysis

Material: 85125320 **SODIUM HYDROXYDE 2,5 N**

Batch : 152718 **Expires end of** 16/09/2016

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	Titre = 2,500 N +/- 0,2 %	2,5012 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

24/09/2015

Signé par:

O. BRINGER , Responsable Qualité