

Certificate of Analysis

Material: 87329290 **ACIDE CHLORHYDRIQUE 0,24 N**

Batch : 152737 **Expires end of** 17/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,2400 N +/- 0,2%	0,2403 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

24/09/2015

Signé par:

O. BRINGER , Responsable Qualité