

Certificate of Analysis

Material: MC3052881000 **Argent nitrate 0,5 N**

Batch : 152774 **Expires end of** 22/09/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,500 N +/- 0,2%	0,499 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/10/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité