

Certificate of Analysis

Material: MC3133631000 **Potassium hydroxyde 0,5 N dans l'éthanol**

Batch : 152958 **Expires end of** 09/04/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,500 N +/- 0,2 %	0,500 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

19/10/2015

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité