

# Certificate of Analysis

---

**Material:** MC3138805000      **TITRIPLEX III 0,025 N**

**Batch :** 153015      **Expires end of** 15/10/2016

| CHARACTERISTICS                         | SPECIFICATIONS              | MESURED VALUES |
|-----------------------------------------|-----------------------------|----------------|
| Titer (20°C; real value 0.2 % accuracy) | Titre = 0,02500 N +/- 0,2 % | 0,02502 N      |

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**  
  
**19/10/2015**

**Signé par:**  
  
O. BRINGER , Responsable Qualité