

Certificate of Analysis

Material: MC3080265000 **Acide chlorhydrique 0,5 N**

Batch : 153274 **Expires end of** 09/11/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,500 N +/- 0,2%	0,500 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

19/11/2015

Signé par:

O. BRINGER , Responsable Qualité