

Certificate of Analysis

Material: MC3023281000 **Potassium hydroxyde N solution alcoolique**

Batch : 153571 **Expires end of** 07/06/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 1,000 N +/- 0,2%	0,998 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

14/12/2015

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité