

Certificate of Analysis

Material: 87642460 **Acide chlorhydrique 3 N**

Batch : 153602 **Expires end of** 08/12/2016

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Titre = 3,000 N +/- 0,2%	3,002 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

14/12/2015

Signé par:

O. BRINGER , Responsable Qualité