

Certificate of Analysis

Material: 85495410 **Potassium Hydroxyde 0,23 N**
Batch : 160453 **Expires end of** 11/02/2017

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	Titre = 0,23 N +/- 0,2%	0,2300 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date
16/02/2016

Signé par:


O. BRINGER , Responsable Qualité