

Certificate of Analysis

Material: 85099320 **Ammonium fer II sulfate en solution à 0,2 mol/l (0,2 N)**
Batch : 2500047 **Expires end of** 08/01/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1996 --> 0,2004 mol/l	0,2000 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

13/01/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité