

Certificate of Analysis

Material: 89375290 **Potassium ferricyanure 0,05 mol/l (0,05 N)**
Batch : 2500083 **en solution aqueuse**
Expires end of 10/01/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,0499 --> 0,0501 mol/l	0,0499 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

13/01/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité