

# Certificate of Analysis

---

**Material:** MC3055040500      **Sodium hydroxyde 0,02 mol/l (0,02 N) en solution aqueuse**

**Batch :** 2500333      **Expires end of** 28/01/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	0,0199 --> 0,0201 mol/l	0,02001 Mol/l

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**03/02/2025**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité