

Certificate of Analysis

Référence : 89372.320

Solution d'acide citrique 3% à pH 8.0

Lot : 2500383

Date de validité

30/01/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Acide Citrique monohydraté Ph Eur Ref VWR 20276 lot 24I124114	30 g +/- 0.5%	conform
Eau HPLC Super Gradient Ref VWR 83650 lot 24I274014	QSP 1000 ml +/- 0.5%	conform
pH (Ajusté à 8.00 avec NaOH Ref 106498 lot MB2012998)	8.00 +/- 0.02	7.98
Filtration	0.22 µ	conform

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Signé par:

A handwritten signature in dark ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité