

Certificate of Analysis

Material: 89715260 **Acide chlorhydrique 0,5 mol/l (0,5 N)**

Batch : 2500481 **Expires end of** 07/02/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,499 --> 0,501 mol/l	0,501 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

10/02/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité