

Certificate of Analysis

Material: 89815290 **Acide citrique en solution à 2,5 M**

Batch : 2500587 **Expires end of** 17/02/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	2,495 --> 2,505 mol/l	11,10 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

24/02/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité