

Certificate of Analysis

Material: 90277180 **EDTA sel disodique 0,001 mol/l (0,002 N) en solution aqueuse**

Batch : 2500665 **Expires end of** 21/02/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,000995-->0,001005 mol/l	0,000999 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

24/02/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité