

Certificate of Analysis

Material: MC3033631000 **Argent nitrate 0.2 mol/l (0.2 N) en solution aqueuse**

Batch : 2500703 **Expires end of** 26/02/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1996 --> 0,2004 mol/l	0,2002 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

03/03/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité